



ACTRIS

CiGas



Germany



France



Finland



Empa

Switzerland



Germany

NOx Data Submission Meeting 2024/02/08



This project receives funding from the European Union's Horizon 2020 research and innovation programme under grant agreements No 654109 and 739530

TOP

1. ACTRIS NO_x in situ data reporting: tasks, responsibilities, and deadlines. (Robert)
2. ACTRIS NO_x in situ data reporting: practical aspects (Yong)
3. Progress on real-time data submission of NO_x data in the CAMS2-21a project (Benni)



Background

Data must be submitted according to the SOP

[Measurement Guideline](#)


and the

[ACTRIS Data Management Plan](#)

until May 31st



Data Management Plan



Defines responsibilities and Data levels:

Until 2022:

Stations submitted corrected level 0 or level 2 data or the TC corrected data

From 2023:

The Stations only submit uncorrected data. The TC does the corrections

NO₂ data measured with Molybdenum converters cannot be accepted

Data Levels

NOx implementation of ACTRIS In Situ data levels:

Level 0: data as provided by instrument, amount fraction and raw counts.

Level 0a: data as provided by instrument, amount fraction and raw counts, flags applied.

Level 1: calibrations applied, original time resolution.

Level 1a: calibrations applied, original time resolution, flags applied.

Level 2: hourly averages, offset correction applied, not sample line corrected.

Level 3: generated directly from level 1, offset correction applied, hourly averages, sample line corrected.



Data Levels


NOx implementation of ACTRIS In Situ data levels:

Level 0a: data as provided by instrument, amount fraction and raw counts, flags applied.

Level 1a: calibrations applied, original time resolution, flags applied.

Level 2: hourly averages, offset correction applied, not sample line corrected.

Level 3: generated directly from level 1, offset correction applied, hourly averages, sample line corrected.



What is needed to produce level 3

Time from entry inlet line to entry of converter: 4.3 s

Duration of stay in converter or bypass line: 0.2 s

Duration of stay in converter: 2.5 s

Converter temperature: 308.15 K

converter efficiency, %

pressure, hPa, Location=inlet, Matrix=instrument

temperature, K, Location=inlet, Matrix=instrument

NO_x Data Level 2

O₃ Data level 1 (calibrated)

Met Data level 0 (raw)




Schedule for Submission of the 2023 Nox Data

March 15th:	NFs:	provide level 0a NO _x Data and level 1a NO _x Data provide Level 1 O ₃ Data / level 0 Met Data
April 09th-11th:	TC / NFs:	QA/QC Workshop
April 30th:	TC / NFs:	QC issues iterated
May 31st:	TC	Level 2 and Level 3 NO _x Data produced



New Data templates

<https://ebas-submit.nilu.no/templates/NOx/all>



EBAS
Data Submission Manual

Controlled Vocabulary Lists ▾

NOx

- [Level 0](#)
- [Level 1](#)
- [Level 2](#)

Data Level	Description	Processing	Used For
0	<ul style="list-style-type: none">Annotated raw dataFormat instrument specificContains all parameters provided by instrument as providedContains all parameters / info needed for processing to final value."Native" time resolution		<ul style="list-style-type: none">Advanced data reportingNear-Real-Time (NRT) data reporting
1	<ul style="list-style-type: none">Data processed to final parameterInvalid data & calibration episodes removed"Native" time resolutionFormat property specificCorrection to standard temperature & pressure cond. (273.15 K, 1013.25 hPa) if necessary		<ul style="list-style-type: none">Advanced data reportingIntercomparisons
1.5	<ul style="list-style-type: none">Data aggregated to hourly averagesAtmospheric variability quantified by standard deviation or percentiles	auto-processed	Near-Real-Time (NRT) data processing
2	<ul style="list-style-type: none">Format property specific	manual quality assurance	Regular, annual data reporting

New Data templates

<https://ebas-submit.nilu.no/templates/NOx/lev0>

Please stick to the “Reduced” Flag list

Group 0: Valid data		
Flag	Validity	Description
000	V	valid data, no flag
Group 1: Exception flags for accepted, irregular data		
Flag	Validity	Description
147	V	Below theoretical detection limit or formal Q/A limit, but a value has been measured and reported and is considered valid
Group 5: Chemical problem		
Flag	Validity	Description
559	V	Unspecified contamination or local influence, but considered valid
Group 6: Mechanical or instrumental problem		
Flag	Validity	Description
686	I	Invalid due to zero check. Used for Level 0.
687	I	Invalid due to span check. Used for Level 0.
699	I	Mechanical problem, unspecified reason
Group 9: Missing flags		
Flag	Validity	Description
999	M	Missing measurement, unspecified reason



New Data templates

<https://ebas-submit.nilu.no/templates/NOx/lev0>

Level 0a: data as provided by instrument, amount fraction and raw counts, flags applied.

NOx measurements by chemiluminescence photometry use either a photolytic or molybdenum converter for transforming NO₂ into NO. It is well established that molybdenum converters convert not only NO₂, but also other reactive nitrogen species. Please note that ACTRIS requires the use of photolytic converters. EMEP and GAW recommends photolytic converters for rural and remote sites. Thus, when submitting data it is essential to specify the type of instrument and converter being used.

This is specified in the metadata element

Instrument type

(
chemiluminescence_photolytic
or
chemiluminescence_molybdenum)



New Data templates

<https://ebas-submit.nilu.no/templates/NOx/lev0>

Zero measurements shall be flagged with flag 686 and the variable status, Status type=zero mode shall be set to indicate whether it was an internal or external zero measurement (0: N/A, 1: internal zero, 2: external zero). See data lines 3 and 4 in the template below for an example for internal and external zero measurements



Comment:

start time	end time	p_inlet	p_det	T_inlet	T_det	cal	zero	NO#	NOc#	NO_sens	cvt_eff	numflag	NO	NO_ac	NO_pr	NO_dl	numflag_NO	NO2	NO2_ac	NO2_pr	NO2_dl	numflag_NO2
0.000000	0.000694	839.400	921.400	299.700	290.210	0	0	258	5200	1.283	45.351	0.0000000000	0.331	0.0800	0.0070	0.005	0.0000000000	10.080	0.0700	0.0060	0.010	0.0000000000
0.000694	0.001385	839.766	921.787	299.750	291.232	0	0	510	4780	1.283	45.351	0.0000000000	0.596	0.0454	0.0066	0.004	0.0000000000	9.765	0.0690	0.0077	0.012	0.0000000000
0.001385	0.002083	839.821	921.833	299.761	291.987	0	1	21	23	1.283	45.351	0.6860000000	0.476	0.0234	0.0087	0.004	0.6860000000	9.832	0.0654	0.0072	0.012	0.6860000000
0.002083	0.002778	839.801	921.788	299.886	292.232	0	2	9	21	1.283	45.351	0.6860000000	0.836	0.0632	0.0077	0.005	0.6860000000	9.922	0.0643	0.0075	0.015	0.6860000000
0.002778	0.003472	839.633	921.654	300.021	292.012	1	0	644	4766	1.283	45.351	0.6870000000	0.839	0.0712	0.0082	0.006	0.6870000000	10.032	0.0704	0.0084	0.012	0.6870000000
0.003472	0.004167	839.601	921.522	299.988	291.987	3	0	523	4802	1.283	45.351	0.6870000000	0.341	0.0673	0.0052	0.004	0.6870000000	9.989	0.0720	0.0060	0.011	0.6870000000
0.004167	0.004861	839.801	921.788	421.122	291.232	0	0	7342	96379	1.283	45.351	0.6990000000	73.832	3.2349	9.7289	8.234	0.6990000000	83.234	32.2724	62.1346	7.234	0.6990000000
0.004861	0.005556	9999.999	9999.999	9999.999	9999.999	99	99	9999	99999	99.999	999.999	0.9990000000	99.999	99.9999	99.9999	99.999	0.9990000000	99.999	99.9999	99.9999	99.999	0.9990000000



New Data templates

<https://ebas-submit.nilu.no/templates/NOx/lev0>

Calibrations shall be flagged with flag 687 and the variable status, Status type=calibration standard shall be set to indicate which calibration gas was used (integer, refers to metadata elements Calibration standard ID and Secondary standard ID)

See data lines 5 and 6 in the template below for an example for calibrations with different standads.

Comment:

start time	end time	p_inlet	p_det	T_inlet	T_det	cal	zero	NO#	NOc#	NO_sens	cvt_eff	numflag	NO	NO_ac	NO_pr	NO_dl	numflag_NO	NO2	NO2_ac	NO2_pr	NO2_dl	numflag_NO2
0.000000	0.000694	839.400	921.400	299.700	290.210	0	0	258	5200	1.283	45.351	0.000000000	0.331	0.0800	0.0070	0.005	0.000000000	10.080	0.0700	0.0060	0.010	0.000000000
0.000694	0.001385	839.766	921.787	299.750	291.232	0	0	510	4780	1.283	45.351	0.000000000	0.596	0.0454	0.0066	0.004	0.000000000	9.765	0.0690	0.0077	0.012	0.000000000
0.001385	0.002083	839.821	921.833	299.761	291.987	0	1	21	23	1.283	45.351	0.686000000	0.476	0.0234	0.0087	0.004	0.686000000	9.832	0.0654	0.0072	0.012	0.686000000
0.002083	0.002778	839.801	921.788	299.886	292.232	0	2	9	21	1.283	45.351	0.686000000	0.836	0.0632	0.0077	0.005	0.686000000	9.922	0.0643	0.0075	0.015	0.686000000
0.002778	0.003472	839.633	921.654	300.021	292.012	1	0	644	4766	1.283	45.351	0.687000000	0.839	0.0712	0.0082	0.006	0.687000000	10.032	0.0704	0.0084	0.012	0.687000000
0.003472	0.004167	839.601	921.522	299.988	291.987	3	0	523	4802	1.283	45.351	0.687000000	0.341	0.0673	0.0052	0.004	0.687000000	9.989	0.0720	0.0060	0.011	0.687000000
0.004167	0.004861	839.801	921.788	421.122	291.232	0	0	7342	96379	1.283	45.351	0.699000000	73.832	3.2349	9.7289	8.234	0.699000000	83.234	32.2724	62.1346	7.234	0.699000000
0.004861	0.005556	9999.999	9999.999	9999.999	9999.999	99	99	9999	99999	99.999	999.999	0.999000000	99.999	99.9999	99.9999	99.999	0.999000000	99.999	99.9999	99.9999	99.999	0.999000000

New Data templates

<https://ebas-submit.nilu.no/templates/NOx/lev0>

```
standard method: SUP=ALUKIS_NOXY_2014
Calibration scale:
Calibration standard ID: "Status calibration standard: 1, Manufacturer: NPL, Batch: A473; Status calibration standard: 2, Manufacturer: Linde, Batch: D736671"
Secondary standard ID: "Status calibration standard: 3, Manufacturer: In House Aluminium cylinder (Luxfer) ... description; Batch: 123; Status calibration standard: 4, Manufacturer: In House Aluminium cylinder (Luxfer) ... description; Batch: 456"
```

Calibration standard ID

Description

Freetext but recommends syntax includes manufacturer and batch number. This refers to the **primary** standard(s) used for calibration (as compared to metadata element [Secondary standard ID](#)). If more than one, please separate with semicolon.

Syntax:

Calibration standard ID: <Freetext>

Example:

Calibration standard ID: "Manufacturer: Linde, Batch: 1234567"



New Data templates

<https://ebas-submit.nilu.no/templates/NOx/lev0>

```
standard method: SUP=ACTRIS_NOXY_2014
Calibration scale:
Calibration standard ID: "Status calibration standard: 1, Manufacturer: NPL, Batch: A473; Status calibration standard: 2, Manufacturer: Linde, Batch: D736671"
Secondary standard ID: "Status calibration standard: 3, Manufacturer: In House Aluminium cylinder (Luxfer) ... description; Batch: 123; Status calibration standard: 4, Manufacturer: In House Aluminium cylinder (Luxfer) ... description; Batch: 456"
```

Secondary standard ID

Description

A secondary standard is derived from a primary standard. It is equivalent to the term "target gas" used by some communities. This metadata field is freetext, but it is recommended that the syntax includes manufacturer and batch number of the secondary standard/target gas. A definition or short description of this standard is also advisable to include. If more than one secondary standard is used, separate by semicolon.

Syntax:

Secondary standard ID: <Freetext>

Example:

Secondary standard ID: "Manufacturer: Linde, Batch: 1234567"

Secondary standard ID: "Manufacturer: Linde, Batch: 1234567; Manufacturer: Linde, Batch: 7654321"



New Data templates

<https://ebas-submit.nilu.no/templates/NOx/lev0>

```

end time of measurement, days from the file reference point
pressure, hPa, Location=inlet, Matrix=instrument
pressure, hPa, Location=detector, Matrix=instrument
temperature, K, Location=inlet, Matrix=instrument
temperature, K, Location=detector, Matrix=instrument
status, no unit, Status type=calibration standard, Matrix=instrument, Comment=See metadata elements "Calibration standard ID" and "Secondary standard ID"
status, no unit, Status type=zero mode, Matrix=instrument, "Comment=0: N/A, 1: internal zero, 2: external zero"
NO #counts, cps
NO_converter #counts, cps
NO_sensitivity, (pmol/mol)/cps
converter_efficiency, %
numflag, no unit
nitrogen monoxide, nmol/mol, Calibration scale=NPL
nitrogen monoxide, nmol/mol, Statistics=expanded uncertainty 2sigma
nitrogen monoxide, nmol/mol, Statistics=precision
nitrogen monoxide, nmol/mol, Statistics=detection limit
numflag, no unit
nitrogen dioxide, nmol/mol, Calibration scale=NPL+GPT
nitrogen dioxide, nmol/mol, Statistics=expanded uncertainty 2sigma
nitrogen dioxide, nmol/mol, Statistics=precision
nitrogen dioxide, nmol/mol, Statistics=detection limit
numflag, no unit
0
    
```

Comment:

start time	end time	p_inlet	p_det	T_inlet	T_det	cal	zero	NO#	NOc#	NO sens	cvt	eff	numflag	NO	NO ac	NO pr	NO dl	numflag NO	NO2	NO2 ac	NO2 pr	NO2 dl	numflag NO2
0.000000	0.000694	839.400	921.400	299.700	290.210	0	0	258	5200	1.283	45.351	0.0000000000	0.331	0.0800	0.0070	0.005	0.0000000000	10.080	0.0700	0.0060	0.010	0.0000000000	
0.000694	0.001385	839.766	921.787	299.750	291.232	0	0	510	4780	1.283	45.351	0.0000000000	0.596	0.0454	0.0066	0.004	0.0000000000	9.765	0.0690	0.0077	0.012	0.0000000000	
0.001385	0.002083	839.821	921.833	299.761	291.987	0	1	21	23	1.283	45.351	0.6860000000	0.476	0.0234	0.0087	0.004	0.6860000000	9.832	0.0654	0.0072	0.012	0.6860000000	
0.002083	0.002778	839.801	921.788	299.886	292.232	0	2	9	21	1.283	45.351	0.6860000000	0.836	0.0632	0.0077	0.005	0.6860000000	9.922	0.0643	0.0075	0.015	0.6860000000	
0.002778	0.003472	839.633	921.654	300.021	292.012	1	0	644	4766	1.283	45.351	0.6870000000	0.839	0.0712	0.0082	0.006	0.6870000000	10.032	0.0704	0.0084	0.012	0.6870000000	
0.003472	0.004167	839.601	921.522	299.988	291.987	3	0	523	4802	1.283	45.351	0.6870000000	0.341	0.0673	0.0052	0.004	0.6870000000	9.989	0.0720	0.0060	0.011	0.6870000000	
0.004167	0.004861	839.801	921.788	421.122	291.232	0	0	7342	96379	1.283	45.351	0.6990000000	73.832	3.2349	9.7289	8.234	0.6990000000	83.234	32.2724	62.1346	7.234	0.6990000000	
0.004861	0.005556	9999.999	9999.999	9999.999	9999.999	99	99	9999	99999	99.999	999.999	0.9990000000	99.999	99.9999	99.9999	99.999	0.9990000000	99.999	99.9999	99.9999	99.999	0.9990000000	



New Data templates

<https://ebas-submit.nilu.no/templates/NOx/lev0>

```

end time of measurement, days from the file reference point
pressure, hPa, Location=inlet, Matrix=instrument
pressure, hPa, Location=detector, Matrix=instrument
temperature, K, Location=inlet, Matrix=instrument
temperature, K, Location=detector, Matrix=instrument
status, no unit, Status type=calibration standard, Matrix=instrument, Comment=See metadata elements "Calibration standard ID" and "Secondary standard ID"
status, no unit, Status type=zero mode, Matrix=instrument, "Comment=0: N/A, 1: internal zero, 2: external zero"
NO #counts, cps
NO_converter #counts, cps
NO_sensitivity, (pmol/mol)/cps
converter_efficiency, %
numflag, no unit
nitrogen monoxide, nmol/mol, Calibration scale=NPL
nitrogen monoxide, nmol/mol, Statistics=expanded uncertainty 2sigma
nitrogen monoxide, nmol/mol, Statistics=precision
nitrogen monoxide, nmol/mol, Statistics=detection limit
numflag, no unit
nitrogen dioxide, nmol/mol, Calibration scale=NPL+GPT
nitrogen dioxide, nmol/mol, Statistics=expanded uncertainty 2sigma
nitrogen dioxide, nmol/mol, Statistics=precision
nitrogen dioxide, nmol/mol, Statistics=detection limit
numflag, no unit
0
    
```

Comment:

start time	end time	p_inlet	p_det	T_inlet	T_det	cal	zero	NO#	NOc#	NO sens	cvt	eff	numflag	NO	NO ac	NO pr	NO dl	numflag NO	NO2	NO2 ac	NO2 pr	NO2 dl	numflag NO2
0.000000	0.000694	839.400	921.400	299.700	290.210	0	0	258	5200	1.283	45.351	0.0000000000	0.331	0.0800	0.0070	0.005	0.0000000000	10.080	0.0700	0.0060	0.010	0.0000000000	
0.000694	0.001385	839.766	921.787	299.750	291.232	0	0	510	4780	1.283	45.351	0.0000000000	0.596	0.0454	0.0066	0.004	0.0000000000	9.765	0.0690	0.0077	0.012	0.0000000000	
0.001385	0.002083	839.821	921.833	299.761	291.987	0	1	21	23	1.283	45.351	0.6860000000	0.476	0.0234	0.0087	0.004	0.6860000000	9.832	0.0654	0.0072	0.012	0.6860000000	
0.002083	0.002778	839.801	921.788	299.886	292.232	0	2	9	21	1.283	45.351	0.6860000000	0.836	0.0632	0.0077	0.005	0.6860000000	9.922	0.0643	0.0075	0.015	0.6860000000	
0.002778	0.003472	839.633	921.654	300.021	292.012	1	0	644	4766	1.283	45.351	0.6870000000	0.839	0.0712	0.0082	0.006	0.6870000000	10.032	0.0704	0.0084	0.012	0.6870000000	
0.003472	0.004167	839.601	921.522	299.988	291.987	3	0	523	4802	1.283	45.351	0.6870000000	0.341	0.0673	0.0052	0.004	0.6870000000	9.989	0.0720	0.0060	0.011	0.6870000000	
0.004167	0.004861	839.801	921.788	421.122	291.232	0	0	7342	96379	1.283	45.351	0.6990000000	73.832	3.2349	9.7289	8.234	0.6990000000	83.234	32.2724	62.1346	7.234	0.6990000000	
0.004861	0.005556	9999.999	9999.999	9999.999	9999.999	99	99	9999	99999	99.999	999.999	0.9990000000	99.999	99.9999	99.9999	99.999	0.9990000000	99.999	99.9999	99.9999	99.999	0.9990000000	



New Data templates

<https://ebas-submit.nilu.no/templates/NOx/lev1>

Level 1a: calibrations applied, original time resolution, flags applied.

- When processing data from lev0 to lev1: Calibration and zero measurements in lev0 shall be used
- All values with **invalid flags** in lev0 will be taken out in lev1 (value = MISSING VALUE, flag = 999)
- All **valid flags** are taken over into lev1.
- NOx shall be calculated from NO and NO₂ measurements.



New Data templates

<https://ebas-submit.nilu.no/templates/NOx/lev1>

Level 1a: calibrations applied, original time resolution, flags applied.



Group 0: Valid data

Flag	Validity	Description
000	V	valid data, no flag

Group 1: Exception flags for accepted, irregular data

Flag	Validity	Description
147	V	Below theoretical detection limit or formal Q/A limit, but a value has been measured and reported and is considered valid

Group 5: Chemical problem

Flag	Validity	Description
559	V	Unspecified contamination or local influence, but considered valid

Group 9: Missing flags

Flag	Validity	Description
999	M	Missing measurement, unspecified reason

New Data templates

<https://ebas-submit.nilu.no/templates/NOx/lev1>

Level 1a: calibrations applied, original time resolution, flags applied.

```
end time of measurement, days from the file reference point  
pressure, hPa, Location=inlet, Matrix=instrument  
temperature, K, Location=inlet, Matrix=instrument  
numflag, no unit  
nitrogen monoxide, nmol/mol, Calibration scale=NPL  
nitrogen monoxide, nmol/mol, Statistics=expanded uncertainty 2sigma  
nitrogen monoxide, nmol/mol, Statistics=precision  
nitrogen monoxide, nmol/mol, Statistics=detection limit  
numflag, no unit  
nitrogen dioxide, nmol/mol, Calibration scale=NPL+GPT  
nitrogen dioxide, nmol/mol, Statistics=expanded uncertainty 2sigma  
nitrogen dioxide, nmol/mol, Statistics=precision  
nitrogen dioxide, nmol/mol, Statistics=detection limit  
numflag, no unit  
NOx, nmol/mol  
NOx, nmol/mol, Statistics=expanded uncertainty 2sigma  
NOx, nmol/mol, Statistics=precision  
NOx, nmol/mol, Statistics=detection limit  
numflag, no unit
```



New Data templates

<https://ebas-submit.nilu.no/templates/NOx/lev1>

Level 0a: data as provided by instrument, amount fraction and raw counts, flags applied..

Comment:

start time	end time	p_inlet	p_det	T_inlet	T_det	cal	zero	NO#	NOc#	NO_sens	cvt_eff	numflag	NO	NO_ac	NO_pr	NO_dl	numflag NO	NO2	NO2_ac	NO2_pr	NO2_dl	numflag NO2
0.000000	0.000694	839.400	921.400	299.700	290.210	0	0	258	5200	1.283	45.351	0.000000000	0.331	0.0800	0.0070	0.005	0.000000000	10.080	0.0700	0.0060	0.010	0.000000000
0.000694	0.001385	839.766	921.787	299.750	291.232	0	0	510	4780	1.283	45.351	0.000000000	0.596	0.0454	0.0066	0.004	0.000000000	9.765	0.0690	0.0077	0.012	0.000000000
0.001385	0.002083	839.821	921.833	299.761	291.987	0	1	21	23	1.283	45.351	0.686000000	0.476	0.0234	0.0087	0.004	0.686000000	9.832	0.0654	0.0072	0.012	0.686000000
0.002083	0.002778	839.801	921.788	299.886	292.232	0	2	9	21	1.283	45.351	0.686000000	0.836	0.0632	0.0077	0.005	0.686000000	9.922	0.0643	0.0075	0.015	0.686000000
0.002778	0.003472	839.633	921.654	300.021	292.012	1	0	644	4766	1.283	45.351	0.687000000	0.839	0.0712	0.0082	0.006	0.687000000	10.032	0.0704	0.0084	0.012	0.687000000
0.003472	0.004167	839.601	921.522	299.988	291.987	3	0	523	4802	1.283	45.351	0.687000000	0.341	0.0673	0.0052	0.004	0.687000000	9.989	0.0720	0.0060	0.011	0.687000000
0.004167	0.004861	839.801	921.788	421.122	291.232	0	0	7342	96379	1.283	45.351	0.699000000	73.832	3.2349	9.7289	8.234	0.699000000	83.234	32.2724	62.1346	7.234	0.699000000
0.004861	0.005556	9999.999	9999.999	9999.999	9999.999	99	99	9999	99999	99.999	999.999	0.999000000	99.999	99.9999	99.9999	99.999	0.999000000	99.999	99.9999	99.9999	99.999	0.999000000



Level 1a: calibrations applied, original time resolution, flags applied.

Comment:

start time	end time	p_inlet	T_inlet	numflag	NO	NO_ac	NO_pr	NO_dl	numflag NO	NO2	NO2_ac	NO2_pr	NO2_dl	numflag NO2	NOx	NOx_ac	NOx_pr	NOx_dl	numflag NOx
0.000000	0.000694	839.400	299.700	0.000000000	0.331	0.0800	0.0070	0.005	0.000000000	10.080	0.0700	0.0060	0.010	0.000000000	10.411	0.1500	0.0125	0.010	0.000000000
0.000694	0.001385	839.766	299.750	0.000000000	0.596	0.0454	0.0066	0.004	0.000000000	9.765	0.0690	0.0077	0.012	0.000000000	11.361	0.1144	0.0142	0.012	0.000000000
0.001385	0.002083	839.821	299.761	0.000000000	99.999	99.9999	99.9999	99.999	0.999000000	99.999	99.9999	99.9999	99.999	0.999000000	999.999	99.9999	99.9999	99.999	0.999000000
0.002083	0.002778	839.801	299.886	0.000000000	99.999	99.9999	99.9999	99.999	0.999000000	99.999	99.9999	99.9999	99.999	0.999000000	999.999	99.9999	99.9999	99.999	0.999000000
0.002778	0.003472	839.633	300.021	0.000000000	99.999	99.9999	99.9999	99.999	0.999000000	99.999	99.9999	99.9999	99.999	0.999000000	999.999	99.9999	99.9999	99.999	0.999000000
0.003472	0.004167	839.601	299.988	0.000000000	99.999	99.9999	99.9999	99.999	0.999000000	99.999	99.9999	99.9999	99.999	0.999000000	999.999	99.9999	99.9999	99.999	0.999000000
0.004167	0.004861	839.801	339.122	0.000000000	99.999	99.9999	99.9999	99.999	0.999000000	99.999	99.9999	99.9999	99.999	0.999000000	999.999	99.9999	99.9999	99.999	0.999000000
0.004861	0.005556	9999.999	9999.999	0.999000000	99.999	99.9999	99.9999	99.999	0.999000000	99.999	99.9999	99.9999	99.999	0.999000000	999.999	99.9999	99.9999	99.999	0.999000000



